

Scientific Paper:

Bioprocess Biosyst Eng., DOI 10.1007/s00449-009-0367-0, 2009

Optical device for parallel online measurement of dissolved oxygen and pH in shake flask cultures

Konstantin Schneider¹, Verena Schütz¹, Gernot Thomas John², Elmar Heinzle¹

¹Biochemical Engineering Institute, Saarland University, 66123 Saarbrücken, Germany

²PreSens Precision Sensing GmbH, 93053 Regensburg, Germany

Abstract:

We describe a new device with parallel optical measurement of dissolved oxygen (DO) and pH in up to nine shake flasks applicable in any conventional shaking incubator. Measurement ranges are 0–500% of air saturation for oxygen and 5.5–8.5 for pH. It was used to characterize growth profiles of different ϵ -lysine producing strains of *Corynebacterium glutamicum*, of *Saccharomyces cerevisiae* and of *Escherichia coli*. Cultures in unbaffled flasks were highly reproducible. Oxygen limitation was indicated online which is particularly important when cultivating fast growing cells as *E. coli*. *C. glutamicum* strains showed distinct characteristic patterns of DO and pH indicating biological events. During the cultivation of *S. cerevisiae* on glucose, fructose and galactose, oxygen uptake rate was determined using the predetermined value of k_{La} . pH measurement was used to determine the minimum buffer requirement for a culture of *C. glutamicum*.

Key-words: Shake flask, Optode, Dissolved oxygen, pH, Online measurement